

测温环外径与温度对照表

产品型号(Type): CTC 019					规格 (Dimensions): $\Phi_{o20} \times \Phi_i 10 \times h 7\text{mm}$, 浅兰 (Light blue)						
适用温度范围(Temperature): 1100 °C ~ 1400 °C					试验条件(Condition): 空气气氛下保温 1 小时						
Φ (mm)	T (°C)	Φ (mm)	T (°C)	Φ (mm)	T (°C)	Φ (mm)	T (°C)	Φ (mm)	T (°C)	Φ (mm)	T (°C)
		19.42	1182	18.92	1242	18.42	1309	17.92	1342	17.42	1372
		19.41	1183	18.91	1244	18.41	1310	17.91	1342	17.41	1372
		19.40	1184	18.90	1246	18.40	1311	17.90	1343	17.40	1373
		19.39	1185	18.89	1247	18.39	1311	17.89	1344	17.39	1373
		19.38	1186	18.88	1249	18.38	1312	17.88	1344	17.38	1374
19.87	1100	19.37	1188	18.87	1251	18.37	1313	17.87	1345	17.37	1375
19.86	1102	19.36	1189	18.86	1253	18.36	1314	17.86	1345	17.36	1376
19.85	1104	19.35	1190	18.85	1255	18.35	1315	17.85	1346	17.35	1376
19.84	1106	19.34	1191	18.84	1256	18.34	1316	17.84	1346	17.34	1377
19.83	1108	19.33	1192	18.83	1258	18.33	1316	17.83	1347	17.33	1377
19.82	1110	19.32	1193	18.82	1260	18.32	1317	17.82	1348	17.32	1378
19.81	1112	19.31	1194	18.81	1262	18.31	1317	17.81	1348	17.31	1379
19.80	1114	19.30	1195	18.80	1263	18.30	1318	17.80	1349	17.30	1380
19.79	1116	19.29	1197	18.79	1265	18.29	1319	17.79	1350	17.29	1381
19.78	1119	19.28	1198	18.78	1267	18.28	1320	17.78	1350	17.28	1381
19.77	1121	19.27	1199	18.77	1268	18.27	1320	17.77	1351	17.27	1382
19.76	1124	19.26	1200	18.76	1270	18.26	1321	17.76	1351	17.26	1382
19.75	1126	19.25	1201	18.75	1272	18.25	1322	17.75	1352	17.25	1383
19.74	1128	19.24	1202	18.74	1273	18.24	1322	17.74	1352	17.24	1384
19.73	1131	19.23	1203	18.73	1275	18.23	1323	17.73	1353	17.23	1385
19.72	1133	19.22	1204	18.72	1276	18.22	1324	17.72	1353	17.22	1385
19.71	1136	19.21	1205	18.71	1278	18.21	1324	17.71	1354	17.21	1386
19.70	1138	19.20	1206	18.70	1280	18.20	1325	17.70	1354	17.20	1387
19.69	1141	19.19	1207	18.69	1281	18.19	1326	17.69	1355	17.19	1388
19.68	1143	19.18	1208	18.68	1282	18.18	1326	17.68	1355	17.18	1389
19.67	1145	19.17	1210	18.67	1284	18.17	1327	17.67	1356	17.17	1390
19.66	1147	19.16	1211	18.66	1285	18.16	1328	17.66	1357	17.16	1390
19.65	1149	19.15	1212	18.65	1286	18.15	1328	17.65	1357	17.15	1391
19.64	1151	19.14	1213	18.64	1288	18.14	1329	17.64	1358	17.14	1392
19.63	1153	19.13	1214	18.63	1289	18.13	1329	17.63	1358	17.13	1393
19.62	1155	19.12	1215	18.62	1290	18.12	1330	17.62	1359	17.12	1394
19.61	1156	19.11	1216	18.61	1291	18.11	1331	17.61	1359	17.11	1394
19.60	1158	19.10	1218	18.60	1292	18.10	1331	17.60	1360	17.10	1395
19.59	1159	19.09	1219	18.59	1293	18.09	1332	17.59	1361	17.09	1396
19.58	1161	19.08	1220	18.58	1294	18.08	1332	17.58	1361	17.08	1397
19.57	1163	19.07	1221	18.57	1295	18.07	1333	17.57	1362	17.07	1398
19.56	1164	19.06	1223	18.56	1296	18.06	1333	17.56	1363	17.06	1398
19.55	1166	19.05	1224	18.55	1297	18.05	1334	17.55	1363	17.05	1399
19.54	1167	19.04	1225	18.54	1298	18.04	1335	17.54	1364	17.04	1400
19.53	1168	19.03	1226	18.53	1299	18.03	1335	17.53	1364		
19.52	1170	19.02	1228	18.52	1300	18.02	1336	17.52	1365		
19.51	1171	19.01	1229	18.51	1301	18.01	1337	17.51	1366		
19.50	1172	19.00	1230	18.50	1302	18.00	1337	17.50	1366		
19.49	1174	18.99	1232	18.49	1303	17.99	1338	17.49	1367		
19.48	1175	18.98	1233	18.48	1304	17.98	1338	17.48	1368		
19.47	1176	18.97	1235	18.47	1305	17.97	1339	17.47	1368		
19.46	1177	18.96	1236	18.46	1306	17.96	1340	17.46	1369		
19.45	1179	18.95	1237	18.45	1307	17.95	1340	17.45	1369		
19.44	1180	18.94	1239	18.44	1307	17.94	1341	17.44	1370		
19.43	1181	18.93	1241	18.43	1308	17.93	1341	17.43	1371		

检验工程师 (Inspector): 黄满连

生产日期 (Date): 2021-10-25